



RapidChek® SELECT Salmonella

RapidChek® SELECT™ Salmonella

Food System, 10001378

Test Kit Includes:

- 2 x 50 test strips in a desiccated canister
- 100 transfer pipettes
- 100 sample tubes
- Package Insert
- 500 g bottle of RapidChek® SELECT™ Primary Media
- 250 mL bottle of RapidChek® SELECT™ Supplement
- 10 g bottle of RapidChek® SELECT™ Secondary Media

Test kit with media system is sufficient for 100 tests and 25 L of media.

RapidChek® SELECT™ Salmonella

Carcass Rinsate System, 10001383

Test Kit Includes:

- 8 x 50 test strips in a desiccated canister
- 400 transfer pipettes
- 400 sample tubes
- Package Insert

- 500 g bottle of RapidChek® SELECT™ Primary Media
- 250 mL bottle of RapidChek® SELECT™ Supplement
- 4 x 10 g bottle of RapidChek® SELECT™ Secondary Media

Test kit with media system is sufficient for 400 tests and 25 L of media.

RapidChek® SELECT™ Salmonella Environmental Test System, 10001386

Test Kit Includes:

- 8 x 50 strips in a desiccated canister
- 400 transfer pipettes
- 400 sample tubes
- Package Insert
- 500 g bottle of RapidChek® SELECT™ Primary Media
- 250 mL bottle of RapidChek® SELECT™ Supplement
- 3 x 10 g bottle of RapidChek® SELECT™ Secondary Media

Test kit with media system is sufficient for 400 environmental samples or 25 L of media.



Also Available:

- RapidChek® SELECT™ Salmonella Enteritidis
- RapidChek® CONFIRM™ Salmonella Enteritidis IMS
- RapidChek® Listeria
- RapidChek® Listeria NextDay™
- RapidChek® E. coli O157 (including H7)
- RapidChek® CONFIRM™ non-O157 STEC IMS



1 Primary Media Preparation

Prepare RapidChek® SELECT™ Salmonella primary media with supplement.*



Sterilized or Distilled Water* (Liters)	RapidChek® SELECT™ Salmonella Primary Media (Grams)	RapidChek® SELECT™ Salmonella Supplement (mL)
3	60 ± 0.2	30
2	40 ± 0.2	20
1	20 ± 0.2	10
0.225 (225 mL)	4.5 ± 0.2	2.25

*The supplement must be stored at 2 – 8 °C when not in use.

**If you are using DI water the rehydrated media must be autoclaved and cooled prior to adding the supplement.

2 Options for Media Preparation

Non-Autoclaved Option:

Add dehydrated media and supplement to sterile water.

Shelf Life:

Three hours within preparation.

Autoclaved Option:


Add dehydrated media to DI water and autoclave. Just prior to use add supplement.

Shelf Life:

Four weeks at 2 – 8 °C or RT without supplement.

2 Primary Sample Enrichment

Add Sample to RapidChek® SELECT™ Salmonella primary media with supplement and incubate 16 to 22 hours at 42 °C.



Sample Type	Volume of Media to add to Stomacher Bag (mL)
Food Sample (25 g)	225 mL
Food Sample (125 g)	1.125 L
Environmental Sponge (7.5 x 4 cm)	60 – 100 mL
Environmental Swab	10 – 20 mL

3 Secondary Media Preparation*

Prepare RapidChek® SELECT™ Salmonella secondary media and aliquot 1 mL of the prepared media into supplied tubes, 1 tube per each enriched sample.




Sterilized water (mL)	RapidChek® SELECT Salmonella Secondary Media (Grams)
100	7.4 ± 0.2
50	3.7 ± 0.1
10	0.74 ± 0.02

*Boiled secondary media can be stored two weeks at 2 – 8 °C.

4 Secondary Sample Enrichment

Transfer 0.1 mL of the enriched sample into the prepared tubes with 1 mL of RapidChek® SELECT™ Salmonella secondary media and incubate for an additional 16 – 22 hours for high burden foods or 6 – 8 hours for low burden foods at 42 °C.



5 Test

Place test strips into tubes with arrows facing down.



6 Read Result

After 10 minutes, read result.
1 Line = Negative
2 Lines = Positive

